

## UK Declaration of Conformity

Herewith, the manufacturer

*IMT Analytics AG*  
Gewerbstrasse 8  
9470 Buchs SG  
Switzerland

declares on sole responsibility that following products:

<b>Description</b>	<b>Type</b>
Flow-Pressure Measurement Device	FlowAnalyser PF-30x (PF-300, PF-301 VAC, PF-302 LOW)
Flow-Pressure Measurement Device	FlowAnalyser PF-300 PRO
Flow-Pressure Measurement Device	FlowAnalyser VenTest (VenTest 800, VenTest 810, VenTest 820)
Flow-Pressure Measurement Device	FlowAnalyser PFC-3000 (PFC-3000A, PFC-3000V, PFC-3000L)

Comply with the applicable UK legislations:

- **The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012**
- **Electrical Equipment (Safety) Regulations 2016**
- **Electromagnetic Compatibility Regulations 2016**

Applying the following standards:

<b>Regulation aspect</b>	<b>EN Standard</b>	<b>UK implementation</b>	<b>Title</b>
<b>Safety</b>	EN 61010-1:2020 + A1:2019	BS EN 61010-1:2010 +A1:2019	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements
<b>EMC</b>	EN IEC 61326-1:2021	BS EN IEC 61326-1:2021	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

This declaration of conformity is issued under the full responsibility of the manufacturer and by the application of the UKCA marking



This declaration is valid for abovementioned products with manufacturing date code 2023-06 (only PF-300 PRO) respectively 2023-08 and later.

**Signed for and on behalf of:**

IMT Analytics AG

**Place and date of issue:**

Buchs SG, 2023-08-10

**Name, function:**

Hans-Jörg Reinhold, Senior Regulatory Affairs Manager

**Signature:**