



RIGEL MEDICAL

GMC-INSTRUMENTS GROUP

What's the maximum amount of data the Rigel 288 can store?

- Max number of asset records 5,000
- Max number of test sequences - 50 (of which 12 are predefined)
- Characters in a test sequence name - 25
- Characters in a User Test name - 200
- Characters in a User Test unit - 25
- Max number of user names held in list - 40
- Max number of entries in the comments string list - 80
- Max number of entries in the AP module name list - 20
- Number of Trace variables (Site, Location etc.) - 8
- Max number of entries in each trace variables list - 40
- Max number of tests per test sequence – 100
- Max of 300 Test Codes depending on size of each test code. (If each code references lots of Test Sequences, then this number will reduce)

If you require more help, please contact us at

<https://www.seaward.com/gb/enquire/>.